



STATE OF WEST VIRGINIA  
DEPARTMENT OF AGRICULTURE

Gus R. Douglass  
Commissioner

1900 Kanawha Boulevard, East  
Charleston, West Virginia 25305-0170  
(304) 558-3550

DL-032796-12

David E. Miller  
Deputy Commissioner

March 27, 1996

Diane Heim  
U.S. Nuclear Regulatory Commission  
101 Manetta Street  
Suite 2900  
Atlanta, GA 30323

Dear Ms. Heim:

On December 28, 1995, I received notification that the West Virginia Department of Agriculture's NRC Material License #47-10656-01 had been terminated. At that time, we had in our possession four electron capture detectors (ECDs), three of which (Vanan Model #02-1972-00, S/Ns A3183, A9784 and A9785) were covered under Vanan's General License #3092-07GL issued by the State of California, Dept. of Health and Services (expiration date April 17, 2002). The fourth detector, a Waters (Tracor) Model #115500-0001 (S/N 5312), was still being covered under our general license, although we were attempting the transfer of coverage to Finnigan Trimetrics, the vendor.

Although the License Transfer Kit for transferring the Waters ECD was requisitioned from Finnigan in November 1995, we did not receive the kit until March 4, 1996. We have adhered to the directions contained therein, taken the necessary wipe tests (results attached), affixed the appropriate labels to the instrument, and completed/submitted the necessary paperwork to affect the transfer. The Waters ECD is now covered under Finnigan's Radioactive Material License #L01186 issued by the Texas Department of Health, Bureau of Radiation Control (expiration date April 30, 1996).

In view of the foregoing, all four ECD units still in our possession are now covered under the vendors' general licenses and termination of our Materials License #47-10656-01 may be finalized. If you have further questions or concerns, I can be reached at (304)-558-2208 FAX (304)-558-3594.

Sincerely,

James A. Campbell, Director  
Regulatory Protection Division

Attachment

A 39 56

# WIPE TEST INSPECTION CERTIFICATE

THIS SECTION TO BE COMPLETED BY TECHNICIAN PERFORMING WIPE TEST

CUSTOMER	West Virginia Department of Agriculture	WIPER TEST PERFORMED BY	Deo Picart	DETECTOR TYPE	NI63	TECH. NUMBER	5312
DIVISION	Regulatory Protection Division	STATE OF NRC LICENSE NUMBER	NRC #47-10656-01	DATE WIPE PERFORMED	3/4/96	EXPIRATION DATE	4/30/99
ADDRESS	4720 Brenda Lane; Bldg #11 Charleston, WV 25312	CITY AND STATE	Georgia	PHONE NUMBER	(304)-558-2208		
ATTENTION	James A. Campbell						

ADDRESS MUST BE SPECIFIC. NO POST OFFICE BOX NUMBERS ALLOWED

## TEST RESULTS

THIS SECTION TO BE COMPLETED BY FINNIGAN CORPORATION

**RADIOACTIVE MATERIAL LICENSE: L01186**  
**ISSUED BY THE STATE OF TEXAS**

THE APPLICANT'S DATA

LESS THAN 0005 UCI (NORMAL LIMIT)

BETWEEN 0005 & 005 UCI (OUTSIDE NORMAL LIMITS)

UCI (CALCULATED)

CUSTOMER NOTIFIED IF OUTSIDE LIMITS

### ELECTRON CAPTURE TEST PARAMETERS

(FINNIGAN CORPORATION ORIGINAL WIPE TEST INSTRUMENT #)

CARRIER	N <sub>2</sub>	FLOW	40 cc/min
SCAVENGER	N <sub>2</sub>	FLOW	20 cc/min
TEMPERATURE	275°C	PARAMETER	GC 107
BACKGROUND			

*W. H. ...* 3/8-96  
DATE

NEXT WIPE TEST DICTATED BY TERMS OF YOUR LICENSE  
(3 YEARS FOR FINNIGAN CORPORATION GENERALLY LICENSED WIPE TESTERS)



TREMETRICS CHROMATOGRAPHY GROUP