

DL-040695_02



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION II
101 MARIETTA STREET, N.W., SUITE 2900
ATLANTA, GEORGIA 30323-0199

April 6, 1995

Riverside Regional Medical Center
ATTN: Leland R. Kirkland, Ph.D.
Radiation Safety Officer
500 T. Clyde Morris Boulevard
Newport News, Virginia 23601-1976

SUBJECT: REQUEST FOR ADDITIONAL INFORMATION CONCERNING THE
APPLICATION DATED December 15, 1994 (Docket No. 030-
10073, Control No. 256261)

Dear Dr. Kirkland:

This letter is in reference to the radioactive materials license application dated December 15, 1994 for the renewal of NRC License No. 45-09001-03.

The following information is needed to complete our review of your application:

1. In your application you indicate that in accordance with the provisions of 10 CFR Part 20, personnel monitoring would not be required. 10 CFR 20.1502 (a) (3), requires that licensees monitor the occupational exposure to radiation of individuals entering a high or very high radiation area. According to your emergency procedures, it is possible that an individual would have to enter a high radiation area in the event that a patient required assistance in leaving the teletherapy room with the source in an unshielded position. Therefore, please confirm that personnel monitoring devices are provided to those individuals who would be expected to respond to an emergency as described above. Also, please note that you are not subject to the dose recording requirements of 10 CFR 20.2106 unless an individual is likely to receive in excess of 10 percent of the limits specified in 10 CFR Part 20, or an individual enters a high or very high radiation area.
2. Please provide a detailed annotated drawing of your facility. Include a description of the composition, position and thickness of shielding that is present in your teletherapy room. Identify adjacent areas to the teletherapy room and point out which areas are considered restricted and unrestricted. Indicate location by building and room number.
3. Please provide a description of your patient viewing system. Also, provide a description of your back-up system or a commitment to suspend all treatments until the viewing system is operable.

Riverside Regional
Medical Center

2

Our review of your radioactive materials license renewal request will continue upon receipt of the requested information. Please provide two copies of your reply and refer to Control No. 256261 in all correspondence regarding this renewal request. If you wish, replies may be transmitted via electronic facsimile (FAX) to (404) 331-7437. You should then mail the original copy of any electronically transmitted documents.

In order to continue review of your renewal application, we request that you submit your response to this letter within thirty (30) days from the date of this letter.

If you have any questions, please call me at (404) 331-3932.

Sincerely,

Michael L. Fuller
Health Physicist
Nuclear Materials Licensing Section

~~RECEIVED~~

~~RECEIVED~~

Miller
12/15/05

JPotter
12/15/05