DL-051398_04

CONVERSATION RECORD

TIME | DATE 5/13/1998

TYPE: ☐ VISIT ☐ CONFERENCE ☒ TELEPHONE ☐ INCOMING ☐ OUTGOING

{ } NEW
{ } RENEW
{ } AMEND
(X) TERM.

Location of Visit/Conference:

NAME OF PERSON(S) CONTACTED OR IN CONTACT WITH YOU: Dallas Boggess

ORGANIZATION (Office, Div., Branch): W. Va. Paving Co.

TELEPHONE NO.: 304 768-9733

SUBJECT: Deficiency Telephone Conversation

CONTROL: 257889

DOCKET: 030-12648

LICENSE NUMBER: 47-17368-01

SUMMARY

(1) need request to terminate the license,
(2) need date for the fax form
(3) final survey (leak test area survey) acceptable for sealed sources.

Spoke with Mr. Boggess 5/15/1998. He confirmed all the above.

ACTION REQUIRED: SEE OVER Y N

terminate license.

DUE DATE

NAME OF PERSON DOCUMENTING CONVERSATION: David J. Collins

SIGNATURE: David J Collins

DATE: 5/15/1998

ACTION TAKEN

SIGNATURE | TITLE | DATE

(FOR LFMS USE)
INFORMATION FROM LTS

BETWEEN:

License Fee Management Branch, ARM
and
Regional Licensing Sections

Program Code: 03121
Status Code: 0
Fee Category: 3P
Exp. Date: 20020731
Fee Comments:

Decom Fin Assur Reqd: N

LICENSE FEE TRANSMITTAL

A. REGION II

1. APPLICATION ATTACHED
Applicant/Licensee: ORDERS & HAYNES PAVING CO., INC.
Received Date: 980413
Docket No: 3012648
Control No.: 257889
License No.: 47-17368-01
Action Type: Termination

2. FEE ATTACHED
Amount: NONE
Check No.:

3. COMMENTS

Signed DIANE HEIM
Date 4/14/98

B. LICENSE FEE MANAGEMENT BRANCH (Check when milestone 03 is entered / ✓ /)

1. Fee Category and Amount: 3P FEE EXEMPT

2. Correct Fee Paid / Application may be processed for: Kim
Amendment ✓
Renewal
License

3. OTHER

Signed Rita Messier
Date 4/21/98

4/20/98
Cop. 3 II
Kim
4/21/98

NRC FORM 314 (6-95) U.S. NUCLEAR REGULATORY COMMISSION

10 CFR 30.36(d)(1)(iv)
10 CFR 40.42(g)(1)(iv)
10 CFR 70.38(d)(1)(iv)

APPROVED BY OMB: NO. 3150-0028 EXPIRES: 06/30/98

ESTIMATED BURDEN PER RESPONSE TO COMPLY WITH THIS MANDATORY INFORMATION COLLECTION REQUEST: 30 MINUTES. THIS SUBMITTAL IS USED BY NRC AS PART OF THE BASIS FOR ITS DETERMINATION THAT THE FACILITY HAS BEEN CLEANED OF RADIOACTIVE MATERIAL BEFORE THE FACILITY IS RELEASED FOR UNRESTRICTED USE. FORWARD COMMENTS REGARDING BURDEN ESTIMATE TO THE INFORMATION AND RECORDS MANAGEMENT BRANCH (T-6 F33), U.S. NUCLEAR REGULATORY COMMISSION, WASHINGTON, DC 20555-0001, AND TO THE PAPERWORK REDUCTION PROJECT (3150-0028), OFFICE OF MANAGEMENT AND BUDGET, WASHINGTON, DC 20503. AN AGENCY MAY NOT CONDUCT OR SPONSOR, AND A PERSON IS NOT REQUIRED TO RESPOND TO, A COLLECTION OF INFORMATION UNLESS IT DISPLAYS A CURRENTLY VALID OMB CONTROL NUMBER.

CERTIFICATE OF DISPOSITION OF MATERIALS

INSTRUCTIONS: ALL ITEMS MUST BE COMPLETED -- PRINT OR TYPE
SEND THE COMPLETED CERTIFICATE TO THE NRC OFFICE SPECIFIED ON THE REVERSE

LICENSEE NAME AND ADDRESS
Orders & Haynes Paving Co
P.O. Box 9488
So Chas W. Va. 25309

LICENSE NUMBER
47-17368-01

LICENSE EXPIRATION DATE
July 2002

A. MATERIALS DATA *(Check one and complete as necessary)*

THE LICENSEE OR ANY INDIVIDUAL EXECUTING THIS CERTIFICATE ON BEHALF OF THE LICENSEE CERTIFIES THAT:
(Check and/or complete the appropriate item(s) below.)

☐ 1. NO MATERIALS HAVE EVER BEEN PROCURED OR POSSESSED BY THE LICENSEE UNDER THIS LICENSE.

OR

☒ 2. ALL ACTIVITIES AUTHORIZED BY THE LICENSE HAVE CEASED AND ALL MATERIALS PROCURED AND/OR POSSESSED BY THE LICENSE NUMBER CITED ABOVE HAVE BEEN DISPOSED OF IN THE FOLLOWING MANNER. *(If additional space is needed, use the reverse side or provide attachments.)*

Describe specific material transfer actions and, if there were radioactive wastes generated in terminating this license, the disposal actions including the disposition of low-level radioactive waste, mixed waste, Greater-than-Class-C waste, and sealed sources, if applicable.

For transfers, specify the date of the transfer, the name of the license recipient, and the recipient's NRC license number or Agreement State name and license number.

3/15/98 Transferred to W. Va. Paving Co.
Lic # 47-17070-01

If materials were disposed of directly by the licensee rather than transferred to another licensee, licensed disposal site or waste contractor, describe the specific disposal procedures *(e.g., decay in storage)*

per Telecon 5/15/1998 with Mr. Boggess

B. OTHER DATA

☒ 1. OUR LICENSE HAS NOT YET EXPIRED; PLEASE TERMINATE IT.

2. A RADIATION SURVEY WAS CONDUCTED BY THE LICENSEE TO CONFIRM THE ABSENCE OF LICENSED RADIOACTIVE MATERIALS AND TO DETERMINE WHETHER ANY CONTAMINATION REMAINS ON THE PREMISES COVERED BY THE LICENSE. *(Check one)*

☐ NO *(Attach explanation)*
☒ YES, THE RESULTS *(Check one)*
☒ ARE ATTACHED, or
☐ WERE FORWARDED TO NRC ON *(Date)*

3. THE PERSON TO BE CONTACTED REGARDING THE INFORMATION PROVIDED ON THIS FORM	NAME	TELEPHONE NUMBER (Include Area Code)
	Dallas Boggess	304-165-4723

4. MAIL ALL FUTURE CORRESPONDENCE RELATING THIS LICENSE TO

W. Va. Paving Co. P.O. Box 544 Dunbar W. Va. 25064

CERTIFYING OFFICIAL

I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

PRINTED NAME AND TITLE	SIGNATURE	DATE
Dallas L. Boggess QC Mgr RSO	Dallas L. Boggess	3/25/98

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECTS. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTIONS.

PRINTED ON RECYCLED PAPER

FAX TRANSMITTAL

To Dallas	From Diane NRC
Dept./Agency W V Paving	Phone # 404-562-4723
Fax # 304-755-7911	Fax #

NSN 7540-01-317-7368 5000-101 GENERAL SERVICES ADMINISTRATION

To: _____ License No. 47-17368-01

Company Orders & Haynes Paving Title _____

Fax No. (____) ____ - ____

FROM: Diane Heim

Nuclear Materials Licensing/Inspection Branch _____.

Division of Nuclear Materials Safety

Fax No. (404) 562-4955 Voice (404) 562-4723

Date 1/28/97

SUBJECT: YOUR REQUEST FOR TRANSFER OR TERMINATION OF NRC LICENSE

Please provide the following information, in addition to the NRC Form 314 (Certificate of Disposition of Materials):

All records considered important to the safe and effective decommissioning of the facility, in accordance with 10 CFR 30.35(g), 40.36(f), §70.25(g) and §72.30(d); and all records concerning public dose and waste disposal, have been transferred to:

☐ 1. Name _____ (Successor)
License No. _____
Address _____
City _____ State _____
Phone (____) ____-____ Fax No. (____) ____-____

☐ 2. US NRC Region II, Nuclear Materials Licensing/Inspection Branch, Division of Nuclear Materials Safety

OR

☑ 3. There is no residual contamination of the facility or environs from licensed materials.
Signature Dallas L. Boggess

Printed Name and Title Dallas L. Boggess GC-Mgr RFO

G:\DNMS\FORMS\Term.fax 3/25/1998 Date

per telcon with Mr. Boggess 5/15/1998

TROXLER

3008 Cornwallis Rd. P.O. Box 12057 Research Triangle Park
North Carolina 27709, U.S.A.

Device - Model # Storage, Serial # Area

Source(s) - Serial # _____, Radionuclide

Serial # _____, Radionuclide

Date of Test: 3-17-98

Please print legibly and firmly. This is your return address label.

- [redacted]
- ORDERS & HAYNES
- P.O. Box 9488 Nitro, W. Va. 25304
-

Your Name DENNIS McCOY

Telephone (304) 768-0007

ORIGINAL COPY

LEAK TEST ANALYSIS

This certifies that the sample accompanying this form has been analyzed using an approved monitoring method that measures both beta/gamma & alpha contamination, and, that the results of this analysis shows the removable activity to be less than 0.005 microcuries

Troxler Electronic Laboratories, Inc.
3008 Cornwallis Rd., P O Box 12057
Research Triangle Park, NC 27709
Tel: (919) 549-8661 Fax: (919) 549-0761
License: NC 032-0182-1

LEAK TEST CERTIFICATE

DEVICE:

Model: STORAGE **Serial No:** AREA **Sample Date:** 3-17-98

LEAK TEST ANALYSIS:

The sample has been analyzed using a counting system capable of measuring beta, gamma, and alpha contamination.

Date of last calibration: 12/15/97

	ALPHA	BETA-GAMMA
Conversion factor (cpm/uCi)	4.60E+05	7.43E+05
Background measurement (cpm)	1	31
Sample measurement (cpm)	0	21
Activity (uCi)	< MDA	< MDA
Min. Detectable Activity (uCi)	7.6E-06	1.1E-05

This certifies that the above leak test results are:

☒ Less than 0.005 uCi (200 Bq) ☐ Greater than 0.005 uCi (200 Bq)

Measurement Date: 3/20/98 Certified by: [illegible]

If greater than 0.005 uCi :

Person Notified _____ Date _____

Phone _____ and/or Fax _____

Rev. 4/97

[illegible]