DL-052395_02

CONVERSATION RECORD

TIME | DATE

TYPE ☐ VISIT ☐ CONFERENCE ☒ TELEPHONE ☐ INCOMING ☒ OUTGOING

Location of Visit/Conference:

NAME OF PERSON(S) CONTACTED OR IN CONTACT WITH YOU	ORGANIZATION (Office, dept., bureau, etc.)	TELEPHONE NO.
GARY PELTIER	GENICOM CORP	949-~~1660~~ 703 1162

License 45-06589-01
Docket 30-06568
Control 256195

SUBJECT: LICENSING DEFICIENCY TELEPHONE CALL

{ } New ☒ Renewal { } Amend

SUMMARY

FAX 301-816-5100

(1) App B - needs new telephone no. for HDO

(2) Rad Safety Program (1) calibration of survey instruments not specified (2) type & range of survey instruments not given (3) number of instruments not given.
Isolate ratemeter MK-5 SN 30204
[illegible]

(3) DFP submitted, response sent by EGW.

(4) Need to run FPA COMPLY program based on total annual possession.

SEE OVER - Yes No

Due Date

ACTION REQUIRED

[illegible]
additional data record [illegible]

NAME OF PERSON DOCUMENTING CONVERSATION	SIGNATURE	DATE
David J. Collins	[signature]	1/23/95

ACTION TAKEN

SIGNATURE	TITLE	DATE

50271-101 CONVERSATION RECORD OPTIONAL FORM 271 (12-76) DEPARTMENT OF DEFENSE

Kr 85 7E-7 uCi/ml - x 0.1 = 7E-8 uCi/ml

calculated max. Kr 30 Ci ÷ 3.15E7 sec./yr = .95 uCi/sec.

yearly @ 2200 cfm = 2200 $\frac{ft^3}{min.} \times \frac{60 min}{hr} \times \frac{360 day}{yr} \times \frac{24 hr}{day} \times \frac{28.32 \times 10^3 ml}{ft^3}$

= 3.23E13 $\frac{ml}{yr}$

= 7E-8 $\frac{uCi}{ml} \times$ 3.23E13 $\frac{ml}{yr}$ = 2.261

alternative uCi/yr: 2.261 $\frac{Ci}{yr}$

[illegible]

[illegible] = [illegible] x 13.13 % of App B limit.