

DL-070199-03

July 1, 1999

**Mark S. Lesser, Chief
Materials Licensing/Inspection Branch 2
Division of Nuclear Materials Safety
United States Nuclear Regulatory Commission
Region II
Atlanta Federal Center
61 Forsyth Street, SW. Suite 23T85
Atlanta, Georgia 30303-3415**

**Subject: Termination of NRC License No.
52-17902-02**

Gentlemen:

This is to acknowledge receipt of your letter of June 3, 1999 and Form 314 "Certificate of Disposition of materials".


As stated in our letter dated May 18, 1999, Caribbean Soil Testing Co., Inc ceased operations in May 1998, therefore we would like to confirm our interest to terminate our license.

Find herewith Form 314 (with one attachment) properly completed.

We like to thanks you for your cooperation in this matter.

Sincerely,

Caribbean Soil Testing Co., Inc.


Alfonso Vazquez

AV

enclosure: NRC Form 314 with attachment

NRC FORM 314
(7-1988)
10 CFR 30.38(c)(1)(iv)
10 CFR 40.42(c)(1)(iv)
10 CFR 70.38(c)(1)(iv)

U.S. NUCLEAR REGULATORY COMMISSION

APPROVED BY OMB NO 3160-0028

EXPIRES 07/31/2001

Estimated burden per response to comply with this mandatory information collection request: 30 minutes. This submittal is used by NRC as part of the basis for its determination that the facility has been cleared of radioactive material before the facility is released for unrestricted use. Forward comments regarding burden estimate to the Records Management Branch (T-6 F33), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001 and to the Paperwork Reduction Project (3150-0028), Office of Management and Budget, Washington, DC 20503. If an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.

CERTIFICATE OF DISPOSITION OF MATERIALS

INSTRUCTIONS: ALL ITEMS MUST BE COMPLETED - PRINT OR TYPE
SEND THE COMPLETED CERTIFICATE TO THE NRC OFFICE SPECIFIED ON THE REVERSE

LICENSEE NAME AND ADDRESS

CARIBBEAN Soil Testing Co., Inc.
P.O. Box 363967
San Juan, P.R. 00936-5967

LICENSE NUMBER

52-17902-02

LICENSE EXPIRATION DATE

2003
October 31, 2003

A. MATERIALS DATA (Check one and complete as necessary)

THE LICENSEE OR ANY INDIVIDUAL EXECUTING THIS CERTIFICATE ON BEHALF OF THE LICENSEE CERTIFIES THAT
(Check and/or complete the appropriate item(s) below)

- 1. NO MATERIALS HAVE EVER BEEN PROCURED OR POSSESSED BY THE LICENSEE UNDER THIS LICENSE
- 2. ALL ACTIVITIES AUTHORIZED BY THE LICENSE HAVE CEASED AND ALL MATERIALS PROCURED AND/OR POSSESSED BY THE LICENSEE UNDER THE LICENSE NUMBER CITED ABOVE HAVE BEEN DISPOSED OF IN THE FOLLOWING MANNER (If additional space is needed, use the reverse side or provide attachments)

Describe specific material transfer actions and, if there were radioactive wastes generated in terminating this license, the disposal actions including the disposition of low-level radioactive waste, mixed waste, Greater-than-Class-C waste, and sealed sources, if applicable.

The nuclear testing machines are those indicated below and were transferred to other NRC Licensed organizations, and there were no radioactive wastes generated in the transfer.

For transfers, specify the date of the transfer, the name of the licensed recipient, and the recipient's NRC license number or Agreement State name and license number.

See Attachment

If materials were disposed of directly by the licensee rather than transferred to another licensee, licensed disposal site or waste contractor describe the specific disposal procedures (e.g., decay in storage)

N/A

B. OTHER DATA

- 1. OUR LICENSE HAS NOT YET EXPIRED. PLEASE TERMINATE IT
- 2. A RADIATION SURVEY WAS CONDUCTED BY THE LICENSEE TO CONFIRM THE ABSENCE OF LICENSED RADIOACTIVE MATERIALS AND TO DETERMINE WHETHER ANY CONTAMINATION REMAINS ON THE PREMISES COVERED BY THE LICENSE

- NO (Attach explanation) see Attachment
- YES, THE RESULTS (Check one)
 - ARE ATTACHED, or
 - WERE FORWARDED TO NRC ON (Date)

- 3. THE PERSON TO BE CONTACTED REGARDING THE INFORMATION PROVIDED ON THIS FORM

NAME
Alfonso Vazquez

TELEPHONE NUMBER
(407) 774-4608

- 4. MAIL ALL FUTURE CORRESPONDENCE REGARDING THIS LICENSE TO

P.O. Box 363967
San Juan, P.R. 00936-3967

AND/OR

Gardenia #3 Box 103
Cape San Uilbge
CAROLINA, P.R. 00971

CERTIFYING OFFICIAL

I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

PRINTED NAME AND TITLE

Alfonso Vazquez, President

SIGNATURE

[Handwritten Signature]

DATE

7/1/99

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECTS. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTIONS.

FILE CERTIFICATES AS FOLLOWS:

IF YOU ARE A DISTRIBUTOR OF EXEMPT PRODUCTS, SEND TO:

DIVISION OF INDUSTRIAL AND MEDICAL NUCLEAR SAFETY
OFFICE OF NUCLEAR MATERIAL SAFETY AND SAFEGUARDS
U.S. NUCLEAR REGULATORY COMMISSION
WASHINGTON, DC 20585-0001

ALL OTHERS, IF YOU ARE LOCATED IN:

CONNECTICUT, DELAWARE, DISTRICT OF COLUMBIA, MAINE,
MARYLAND, MASSACHUSETTS, NEW HAMPSHIRE, NEW
JERSEY, NEW YORK, PENNSYLVANIA, RHODE ISLAND, OR
VERMONT, SEND APPLICATIONS TO:

LICENSING ASSISTANCE SECTION
NUCLEAR MATERIALS SAFETY BRANCH
U.S. NUCLEAR REGULATORY COMMISSION, REGION I
475 ALLENDALE ROAD
KING OF PRUSSIA, PA 19406-1415

ALABAMA, FLORIDA, GEORGIA, KENTUCKY, MISSISSIPPI,
NORTH CAROLINA, PUERTO RICO, SOUTH CAROLINA,
TENNESSEE, VIRGINIA, VIRGIN ISLANDS, OR WEST VIRGINIA,
SEND APPLICATIONS TO:

NUCLEAR MATERIALS SAFETY SECTION
U.S. NUCLEAR REGULATORY COMMISSION, REGION II
ATLANTA FEDERAL CENTER, SUITE 23T85
61 FORSYTH STREET, SW
ATLANTA, GA 30303-3415

IF YOU ARE LOCATED IN:

ILLINOIS, INDIANA, IOWA, MICHIGAN, MINNESOTA, MISSOURI,
OHIO, OR WISCONSIN, SEND APPLICATIONS TO:

MATERIALS LICENSING SECTION
U.S. NUCLEAR REGULATORY COMMISSION, REGION III
801 WARRENVILLE ROAD
LISLE, IL 60532-4351

ALASKA, ARIZONA, ARKANSAS, CALIFORNIA, COLORADO,
HAWAII, IDAHO, KANSAS, LOUISIANA, MONTANA, NEBRASKA,
NEVADA, NEW MEXICO, NORTH DAKOTA, OKLAHOMA,
OREGON, PACIFIC TRUST TERRITORIES, SOUTH DAKOTA,
TEXAS, UTAH, WASHINGTON, OR WYOMING,
SEND APPLICATIONS TO:

MATERIAL RADIATION PROTECTION SECTION
U.S. NUCLEAR REGULATORY COMMISSION, REGION IV
611 RYAN PLAZA DRIVE, SUITE 400
ARLINGTON, TX 76011-8084

ATTACHMENT

CPN Nuclear Machines Identification		Recipient Name	Recipient License Number	Date of Transfer
1	M 18062335	Suelos Inc.	52-19193-01	7/98
2.	M 10073367	Brainard-Kilman	GA 318-1	1/90
3.*	M 19012592	Caribbean Soil Testing (Mayaguez Branch)	52-21368-01	6/83
		Western Soils Inc.	52-21368-01	early 98
4.	M 13115143	Geoconsult Inc.	52-23751	4/96
5.	M 18052313	Geoconsult Inc.	52-23751	2/97
6.	MD 0019312	Suelos Inc.	52-19193-01	7/98

* I will like to add that Nuclear Gauge M 19012592 was bought under our license and immediately transferred to our Mayaguez Branch office in 1983, which later obtained their own license (52-21368-01). This gauge became part of their inventory and custody. Therefore, the leak test and other records for that gauge shall be in the possession of the licensee number 52-21368-01. Afterwards, this license was apparently assigned to Western Soils, Inc.

In relation to the required final Radiation survey, I wish to express again that no survey has been performed since the Building used during our operations are now own by Suelos Inc. which use the same storage area for their nuclear gauges.

0



Advanced Instrumentation for Density & Moisture Testing of Soils & Pavements

CPN International, Inc.
1300 Wood Street
Houston, Texas 77002
Phone: (713) 261-0225
Fax: (713) 261-0226

CERTIFICATE OF LEAK TEST

CARIBBEAN TESTING CO. INC
ATTN: ALFANSO VASQUEZ
GPO BOX 362967
SAN JUAN, PR 00936

Customer No. 3100209
Date Sample Collected 12/22/97
Test Number 22063

This is to certify that the leak test on the indicated source (s) was conducted on 12/22/97 and the results shown accurately represent the level of removable contamination

Model MC-1 Serial Number M1592337

Source	Activity	Removable Contamination (in microCuries)
Am-241	50 mCi	000032
Cs-137	10 mCi	000031

* 0.005 microCuries (185 Bq) or greater constitutes a leaking source.
This Source May Remain in use

Reviewed by Leak Testing Quality Control

4

2830 Howe Street
Marinez, California 94553 USA
Telephone 510-228-9770
Fax: 510-228-3183



**BOART LONGYEAR
CPN**

CERTIFICATE OF LEAK TEST

CARIBBEAN TESTING CO, INC.

Customer Number: 3160200.

Date sample collected:

January 27, 1995

GPO BOX 367021
SAN JUAN, PR 00913

Test Number : 12105
RDC # : 164599

(3)

This is to certify that the leak test on the indicated source(s) was counted on February 7, 1995 and the results shown accurately represent the level of removable contamination.

Model : MC-1 Serial Number : M13115143


Source	Activity	* Removable Contamination +/- Counter Error
-----	-----	-----
Am241/Be	50 mCi	<0.000001 microcuries (< 1 Bq)
Cs137	10 mCi	<0.00001 microcuries (< 1 Bq)

* 0.005 microcuries (185 Bq) or greater constitutes a leaking source.

Reviewed by :
Sean Reilly
Assistant RSO

5

Boart Longyear Company
2830 Howe Road
Martinez, California 94553 USA
Telephone: (510) 228-9770
Fax: (510) 228-3183



BOART LONGYEAR
CPN

CERTIFICATE OF LEAK TEST

CARIBBEAN SOIL TESTING CO., INC.
ATTN: ALFONSO VASQUEZ
P.O. BOX 363967
SAN JUAN P.R. 00936

Customer No: 1111100
Date Sample Collected: 10/30/96
Test Number: 18551

This is to certify that the leak test on the indicated source (s) was counted on 11/06/96 and the results shown accurately represent the level of removable contamination.

Model: MC-1

Serial Number: M18052313

<u>Source</u>	<u>Activity</u>	<u>* Removable Contamination (in microCuries)</u>
AmBe 241	50 mCi	0
Cs 137	10 mCi	.0000171

* 0.005 microcuries (185 Bq) or greater constitutes a leaking source.

This Source : May Remain in use.

Reviewed by: Leak Testing Quality Control

6



Advanced Instrumentation for Density & Moisture Testing of Soils & Pavements

CPN International, Inc.
14000 W. 45th St.
Tampa, FL 33604
Phone: (813) 225-9770
Fax: (813) 225-9141

CERTIFICATE OF LEAK TEST

CARIBBEAN TESTING CO, INC
ATTN: ALFANSO VASQUEZ
P.O. BOX 35395
SAN JUAN, PR 00936

Customer No: 3160200
Date Sample Collected: 12/22/97
Test Number: 22062

This is to certify that the leak test on the indicated source (s) was counted on 12/22/97 and the results shown accurately represent the level of removable contamination.

Model: MC-1DR Serial Number: M09019312

Source	Activity	Removable Contamination (in microCuries)
AmBe 241	50 mCi	.000032
Cs 137	10 mCi	.000031

* 0.005 microcuries (185 Bq) or greater constitutes a leaking source.

This Source May Remain in use

Reviewed by: Leak Testing Quality Control